



CERTIFICATE OF ACCREDITATION

This is to certify that

ABIOLABS LTD

Calibration Laboratory No. C003

is accredited by the ***Mauritius Accreditation Service (MAURITAS)***
for the following Calibration field:

MASS

as per scope of schedule of accreditation

**THIS LABORATORY MEETS THE REQUIREMENTS OF ISO/IEC
17025:2005**

This accreditation demonstrates technical competency for a defined scope and the operation of a laboratory quality management system and shall remain in force subject to continuing compliance with MAURITAS accreditation criteria, ISO/IEC 17025:2005 and any further requirements specified by MAURITAS

Issue Date: 14 August 2018

Director of MAURITAS

This certificate is valid only when accompanied by its schedule of Accreditation.



**Schedule of Accreditation
Laboratory No C003**

Permanent Address of laboratory:

Abiolabs Ltd
6th Floor, Maeva Tower
Silicon Avenue, Ebene Business Park
72201 Reduit

Tel No.: (230) 466 3054

Fax No.: (230) 466 6770

E-mail: jcautres@abiolabs.com

Postal Address:

Abiolabs Ltd
6th Floor, Maeva Tower
Silicon Avenue, Ebene Business Park
72201 Reduit

Technical Signatories:

Mrs. Julie Cautres

Issue No: 01

Expiry Date: 13 August 2022

	<i>Measured Quantity of Type of Gauge or Instrument</i>	<i>Reference to standardized procedure</i>	<i>Range of Measured Quantity</i>	<i>Calibration and Measurement Capabilities Expressed as an Uncertainty (\pm)</i>
I.	Mass			
1.	Weighing Instruments	WI 36-027	0g – 100g >100g – 620g >620g – 4000g >4000g – 15000g	0.001 % 0.001 % 0.0035 % 0.0035 %

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95 %.

Issued by the Mauritius Accreditation Service (MAURITAS)

Date: 14 August 2018

.....
Director of MAURITAS