



CERTIFICATE OF ACCREDITATION

This is to certify that

***LABORATOIRE MEDICAL DE GOODLANDS
OF BIOSANTE LTEE***

Testing Laboratory No.: T020

is accredited by the ***Mauritius Accreditation Service (MAURITAS)***
for the following field:

MEDICAL

as per scope of schedule of accreditation

THIS LABORATORY MEETS THE REQUIREMENTS OF ISO 15189

This accreditation demonstrates technical competency for a defined scope and the operation of a laboratory quality management system and shall remain in force subject to continuing compliance with MAURITAS accreditation criteria, ISO 15189:2012 and any further requirements specified by MAURITAS

Issue Date: 24 July 2017

Director of MAURITAS

This certificate is valid only when accompanied by its schedule of Accreditation.



**Schedule of Accreditation
Laboratory No.: T020**

Permanent Address of laboratory:

Laboratoire Medical de Goodlands
Royal Road, near Tamil Temple
Goodlands

Tel No.: (230) 283 5114

Fax No. : (230) 283 4000

E-mail: jhoomuck@biosante.mu

Postal Address:

Laboratoire Medical de Goodlands
Royal Road, near Tamil Temple
Goodlands

Issue No: 03

Expiry Date: 04 June 2018

	<i>Discipline/ Sample Type</i>	<i>Types of tests/Properties Measured Range of Measurement</i>	<i>Equipment/Method/Technique</i>
I.	Biochemistry		
1.	Plasma/Serum	Alanine Transaminase (ALT)	PICTUS 400 /IFCC / Spectrophotometry
		Aspartate Aminotransferase (AST)	PICTUS 400 / IFCC / Spectrophotometry
		Cholesterol	PICTUS 400 / CHOD-PAP / Spectrophotometry
		Creatinine	PICTUS 400 / MOD, JAFFE / Spectrophotometry
		Gamma Glutamyl Transpeptidase (GGT)	PICTUS 400 / IFCC, SZASZ / Spectrophotometry
		Glucose	PICTUS 400 / GOD-PAP / Spectrophotometry
		Protein Total	PICTUS 400 / BIURET / Spectrophotometry
		Triglycerides	PICTUS 400 / GPO-PAP / Spectrophotometry

		Urea	PICTUS 400 / UREASE / Spectrophotometry
		Uric Acid	PICTUS 400 / URICASE / Spectrophotometry
II.	<i>Haematology</i>		
1.	Blood	Haemoglobin	SYSMEX XS 500i/Sulpholyser.Haemoglobin Method
		Platelets count	SYSMEX XS 500i / RBC Detector Hydro Dynamic Focusing Method
		Red Cells Count	SYSMEX XS 500i / RBC Detector Hydro Dynamic Focusing Method
		Total Leucocyte Count	SYSMEX XS 500i / Flow Cytometry
III.	<i>Immunology</i>		
1.	Serum	Rheumatoid Arthritis Latex	Manual/Latex Agglutination Slide Test

Issued by the Mauritius Accreditation Service (MAURITAS)

Date: 24 July 2017

.....
Director of MAURITAS